

BenchPro, Inc. Airborne Particle Count Record

Area: _____
 Technician: Stefan Hocom
 Testing Date: 2/15/2013

Test: _____
 Sample Volume: 1 ft³

		Cumulative Counts/ft ³					Cumulative Counts/m ³				
Location #	Time	0.1µm	0.2µm	0.3µm	0.5µm	1.0µm	0.1µm	0.2µm	0.3µm	0.5µm	1.0µm
1	13:27:25	0	0	0	0	0	0	0	0	0	0
2	13:28:26	0	0	0	0	0	0	0	0	0	0
3	13:29:27	0	0	0	0	0	0	0	0	0	0

Remarks : Passed ISO Class 3 @ 0.1 micron

BenchPro, Inc. Airborne Particle Count Record

Area: _____
 Technician: Test Chamber
 Testing Date: Stefan Hocom
2/15/2013

Test: _____
 Sample Volume: Static Test 1
1 ft³

		Cumulative Counts/ft ³					Cumulative Counts/m ³				
Location #	Time	0.1µm	0.2µm	0.3µm	0.5µm	1.0µm	0.1µm	0.2µm	0.3µm	0.5µm	1.0µm
1	13:33:30	17	2	0	0	0	600	71	0	0	0
2	13:34:40	10	5	1	0	0	353	177	35	0	0
3	13:37:00	13	4	2	1	1	459	141	71	35	35

Remarks : Passed ISO Class 3 @ 0.1 micron

BenchPro, Inc. Airborne Particle Count Record

Area: _____
 Technician: Test Chamber
 Stefan Hocom
 Testing Date: 2/15/2013

Test: _____
 Sample Volume: Walking Test
 1 ft³

		Cumulative Counts/ft ³					Cumulative Counts/m ³				
Location #	Time	0.1μm	0.2μm	0.3μm	0.5μm	1.0μm	0.1μm	0.2μm	0.3μm	0.5μm	1.0μm
1	13:39:00	123	100	90	73	66	4342	3530	3177	2577	2330
2	13:40:10	64	47	42	37	32	2259	1659	1483	1306	1130
3	13:41:20	97	73	64	57	51	3424	2577	2259	2012	1800

Remarks : Passed ISO Class 5 @ 0.5 micron

BenchPro, Inc. Airborne Particle Count Record

Chair: WAT-VCR
 Technician: Stefan Hocom
 Testing Date: 2/15/2013

Test: Static 2
 Sample Volume: 1 ft³

Location #	Time	Cumulative Counts/ft ³					Cumulative Counts/m ³				
		0.1µm	0.2µm	0.3µm	0.5µm	1.0µm	0.1µm	0.2µm	0.3µm	0.5µm	1.0µm
1	13:44:44	0	0	0	0	0	0	0	0	0	0
2	13:46:23	0	0	0	0	0	0	0	0	0	0
3	13:48:03	0	0	0	0	0	0	0	0	0	0
4	13:49:48	0	0	0	0	0	0	0	0	0	0

Remarks : Indicates ISO class 3 compliance at 0.1 micron

BenchPro, Inc. Airborne Particle Count Record

Chair: WAT-VCR
 Technician: Stefan Hocom
 Testing Date: 2/15/2013

Test: Static 3
 Sample Volume: 1 ft³

Location #	Time	Cumulative Counts/ft ³					Cumulative Counts/m ³					
		0.1µm	0.2µm	0.3µm	0.5µm	1.0µm	0.1µm	0.2µm	0.3µm	0.5µm	1.0µm	
1	14:03:55	21	14	8	3	3	741	494	282	106	106	
2	14:05:05	8	3	3	2	2	282	106	106	71	71	
3	14:06:15	11	7	6	3	2	388	247	212	106	71	
4	14:07:25	8	7	6	2	1	282	247	212	71	35	

Remarks : Indicates ISO class 4 compliance at 0.1 micron

BenchPro, Inc.

Airborne Particle Count Record

Chair: WAT-VCR
 Technician: Stefan Hocom
 Testing Date: 2/15/2013

Test: _____
 Sample Volume: 1 ft³

Location #	Time	Cumulative Counts/ft ³					Cumulative Counts/m ³				
		0.1µm	0.2µm	0.3µm	0.5µm	1.0µm	0.1µm	0.2µm	0.3µm	0.5µm	1.0µm
1	13:53:16	571	462	378	268	199	20156	16309	13343	9460	7025
2	13:54:58	1484	1212	993	710	515	52385	42784	35053	25063	18180
3	13:56:17	1021	825	668	450	317	36041	29123	23580	15885	11190
4	13:58:31	2265	1882	1555	1073	788	79955	66435	54892	37877	27816

Remarks : Indicates ISO class 6 compliance at 0.5 micron when control concentrations are removed (walking test).

BenchPro, Inc. Airborne Particle Count Record

Chair: WAT-VCR
 Technician: Stefan Hocom
 Testing Date: 2/15/2013

Test: Mimic
 Sample Volume: 1 ft³

		Cumulative Counts/ft ³					Cumulative Counts/m ³				
Location #	Time	0.1µm	0.2µm	0.3µm	0.5µm	1.0µm	0.1µm	0.2µm	0.3µm	0.5µm	1.0µm
1	14:11:20	33	26	19	12	11	1165	918	671	424	388
2	14:12:40	40	23	15	9	7	1412	812	530	318	247
3	14:14:00	22	13	7	2	1	777	459	247	71	35
4	14:15:20	19	7	5	3	0	671	247	177	106	0

Remarks : Indicates ISO class 3 compliance at 0.1 micron when control concentrations are removed (walking test).

BenchPro, Inc.

Airborne Particle Count Record

Chair: WAT-VCR
 Technician: Stefan Hocom
 Testing Date: 2/15/2013

Test: _____
 Sample Volume: 1 ft³

Location #	Time	Cumulative Counts/ft ³					Cumulative Counts/m ³				
		0.1µm	0.2µm	0.3µm	0.5µm	1.0µm	0.1µm	0.2µm	0.3µm	0.5µm	1.0µm
1	14:33:14	637	367	265	194	141	22486	12955	9355	6848	4977
2	14:34:29	3536	2938	1286	858	588	124821	103711	45396	30287	20756
3	14:35:44	1243	978	704	499	355	43878	34523	24851	17615	12532
4	14:36:58	1054	726	595	400	289	37206	25628	21004	14120	10202

Remarks : Indicates ISO class 6 compliance at 0.5 micron.

BenchPro, Inc.

Airborne Particle Count Record

Chair: WAT-VCR
 Technician: Stefan Hocom
 Testing Date: 2/15/2013

Test: _____
 Sample Volume: 1 ft³

		Cumulative Counts/ft ³					Cumulative Counts/m ³				
Location #	Time	0.1µm	0.2µm	0.3µm	0.5µm	1.0µm	0.1µm	0.2µm	0.3µm	0.5µm	1.0µm
1	14:19:01	0	0	0	0	0	0	0	0	0	0
	14:19:07	337	286	225	153	120	11896	10096	7943	5401	4236
	14:19:13	26	20	17	12	6	918	706	600	424	212
	14:19:19	24	13	8	5	4	847	459	282	177	141
	14:19:25	2	1	1	0	0	71	35	35	0	0
	14:19:31	4	2	1	0	0	141	71	35	0	0
	14:19:37	4	1	0	0	0	141	35	0	0	0
	14:19:43	2	2	1	1	0	71	71	35	35	0
	14:19:49	1	0	0	0	0	35	0	0	0	0
	14:19:55	1	0	0	0	0	35	0	0	0	0
	14:20:01	0	0	0	0	0	0	0	0	0	0

Remarks : _____ Recovered.

